Planmeca ProOne®

Quick guide

Panorama exposure

Preparation

Select support. ❶ Bite piece ❷ Chin cup ❸ Chin rest	

Select program.	Basic program (Standard)	*Traditional beam path and angles.*
	Interproximal panoramic program (Advanced program, optional)	*Used in the detection of caries.*
	Orthogonal panoramic program (Advanced program, optional)	*X-ray beam perpendicular to the jaw. Practical for examination of the gums.*
	Panoramic program for Bitewing (Bitewing, optional)	*Produces images similar to Bitewing images.*

Select patient size.	

Patient positioning

❶ Position the patient's head so that the Frankfort plane is aligned with the Frankfort plane light beam. ❷ Position the patient's head so that the midsagittal plane is aligned with the midsagittal plane light beam. ❸ Position the patient so that the apices of the upper central incisors are within the unit's image layer (focal trough).	

Exposure

Press the exposure button and hold down for the duration of the exposure.	

30003299-1/9.2016/en

PLANMECA